

Manual of Petroleum Measurement Standards Chapter 22.3

Testing Protocol for Flare Gas Metering

FIRST EDITION, AUGUST 2015



AMERICAN PETROLEUM INSTITUTE

Special Notes

API publications necessarily address problems of a general nature. With respect to particular circumstances, local, state, and federal laws and regulations should be reviewed.

Neither API nor any of API's employees, subcontractors, consultants, committees, or other assignees make any warranty or representation, either express or implied, with respect to the accuracy, completeness, or usefulness of the information contained herein, or assume any liability or responsibility for any use, or the results of such use, of any information or process disclosed in this publication. Neither API nor any of API's employees, subcontractors, consultants, or other assignees represent that use of this publication would not infringe upon privately owned rights.

API publications may be used by anyone desiring to do so. Every effort has been made by the Institute to assure the accuracy and reliability of the data contained in them; however, the Institute makes no representation, warranty, or guarantee in connection with this publication and hereby expressly disclaims any liability or responsibility for loss or damage resulting from its use or for the violation of any authorities having jurisdiction with which this publication may conflict.

API publications are published to facilitate the broad availability of proven, sound engineering and operating practices. These publications are not intended to obviate the need for applying sound engineering judgment regarding when and where these publications should be utilized. The formulation and publication of API publications is not intended in any way to inhibit anyone from using any other practices.

Any manufacturer marking equipment or materials in conformance with the marking requirements of an API standard is solely responsible for complying with all the applicable requirements of that standard. API does not represent, warrant, or guarantee that such products do in fact conform to the applicable API standard.

Users of this Standard should not rely exclusively on the information contained in this document. Sound business, scientific, engineering, and safety judgment should be used in employing the information contained herein.

All rights reserved. No part of this work may be reproduced, translated, stored in a retrieval system, or transmitted by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from the publisher. Contact the Publisher, API Publishing Services, 1220 L Street, NW, Washington, DC 20005.

Copyright © 2015 American Petroleum Institute

Foreword

Nothing contained in any API publication is to be construed as granting any right, by implication or otherwise, for the manufacture, sale, or use of any method, apparatus, or product covered by letters patent. Neither should anything contained in the publication be construed as insuring anyone against liability for infringement of letters patent.

Shall: As used in a standard, “shall” denotes a minimum requirement in order to conform to the specification.

Should: As used in a standard, “should” denotes a recommendation or that which is advised but not required in order to conform to the specification.

This document was produced under API standardization procedures that ensure appropriate notification and participation in the developmental process and is designated as an API standard. Questions concerning the interpretation of the content of this publication or comments and questions concerning the procedures under which this publication was developed should be directed in writing to the Director of Standards, American Petroleum Institute, 1220 L Street, NW, Washington, DC 20005. Requests for permission to reproduce or translate all or any part of the material published herein should also be addressed to the director.

Generally, API standards are reviewed and revised, reaffirmed, or withdrawn at least every five years. A one-time extension of up to two years may be added to this review cycle. Status of the publication can be ascertained from the API Standards Department, telephone (202) 682-8000. A catalog of API publications and materials is published annually by API, 1220 L Street, NW, Washington, DC 20005.

Suggested revisions are invited and should be submitted to the Standards Department, API, 1220 L Street, NW, Washington, DC 20005, standards@api.org.

Contents

	Page
1	Scope 1
2	Normative References 1
3	Terms and Definitions 1
4	Field of Application 2
5	Parameter Variations Affecting Device Performance 3
6	Mandatory Tests 3
6.1	Test Conditions 3
6.2	Test Installation 4
6.3	Test Results 8
6.4	Baseline Testing 8
6.5	Installation Effects Testing 8
6.6	Special Testing 10
6.7	Testing Procedure 10
7	Test Facility Requirements 10
7.1	Audit Process 10
7.2	Lab/Facility Qualification 11
8	Uncertainty Analysis and Calculation 11
8.1	Test Facility Uncertainty 11
8.2	Device Uncertainty 11
8.3	How to Calculate Uncertainty 11
8.4	Presentation of Test Report Uncertainty 11
9	Test Report 11
9.1	General 11
9.2	Test Facility Information 11
9.3	Test Meter Information 12
9.4	Description of the Full Test Matrix And Results 12
9.5	Sample Meter Test Reporting Form 13
	Annex A (informative) Samples for Summary and Detailed Data Reports 14
	Bibliography 21
Figures	
1	Test Program Piping Configuration 7
2	Elbow in Plane Piping Layout with a USM in its Preferred Position 9
3	Elbow out of Plane Piping Layout with a USM in its Preferred Position 9
X	Picture of Meter Under Test 16
Y	Baseline Meter Test Piping Arrangement (example) 18
Z	Baseline Meter Test Piping (example) 18
AA	Baseline Meter Test Result Summary (example) 19
AB	Installation Effect Test Summary—2 Elbows Out of Plane (example) 20
Tables	
X	Reference Test Equipment Summary 17
Y	Baseline Meter Test Summary (example) 19

Introduction

This document defines the testing and reporting protocols for a flare gas meter within a Flare Flow Meter System (FFMS). This protocol is designed to supply industry with a comparable description of the capabilities of flare gas meters for gaseous fluid flow measurement. The following are objectives of this Testing Protocol.

- Provide information about relative performance characteristics of the flare gas meter under a standardized testing protocol over an expected operating range with a standardized reporting format. (See 6.1 for determination of the expected operating range.)
- Facilitate both the understanding and the introduction of new technologies.
- Provide a standardized vehicle for manufacturers to validate the meter's performance.
- Quantify the uncertainty of these devices and define the operating and installation conditions for which the stated uncertainties apply.

To accomplish these objectives, in addition to the base testing requirements defined in this document, the end user should define any application specific requirements not covered by the base testing requirements such as the following.

- Expected operating range of the meter where this exceeds the base testing requirement operating range.
- Upstream flow disturbances not addressed by the base testing requirements.
- Criteria needed to comply with regulatory requirements.

When specifying additional testing requirements, the end user should insure that all parameters affecting the meter's performance are identified, but not limited to, items described in Section 6.

The facility selected to perform the tests has to be able to correlate the test fluid to the fluids to be measured at the expected process conditions.

Examples of flare gas meters covered in this standard include, but are not limited to differential pressure flow meters, optical meters, thermal flow meters, ultrasonic flare meters, vortex shedding meters.

If a meter has been tested under another API *MPMS* Chapter 22 testing protocol that covers the entire expected operating range for this application, testing the influence parameters listed in Section 5 and the ranges defined in 6.4, then additional testing is optional. Reporting and testing protocols for test facilities are included to ensure that the performance characteristics of each meter are compared with identical conditions as set forth in this standard.

Testing Protocol for Flare Gas Metering

1 Scope

The scope of the standard is to describe a testing protocol for flare gas meters. This includes a discussion of the testing to be performed, how the test data should be analyzed, and how an uncertainty is determined from the testing of the meter. The scope does not include the general guidelines to flare gas metering that are covered under *MPMS* Chapter 14.10.

2 Normative References

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

API Manual of Petroleum Measurement Standards (MPMS), Chapter 14.3 Part 1, *Concentric, Square-Edged Orifice Meters—Part 1: General Equations and Uncertainty Guidelines*

ISO/IEC Guide 98-3:2008¹, *Uncertainty of Measurement—Part 3: Guide to the Expression Uncertainty in Measurement (GUM: 1995)*

ISO/IEC 17025; 2005, *General requirements for the competence of testing and calibration laboratories*

3 Terms and Definitions

For the purposes of this document, the following definitions apply.

3.1

accuracy

The ability of a measurement instrument to indicate values closely approximating the true value of the quantity measured.

3.2

adiabatic expansion

Expansion of gas that occurs without loss or gain of heat.

3.3

calibration

The process or procedure of adjusting an instrument, such as a meter, so that its indication or registration is in satisfactorily close agreement with a reference standard.

3.4

D

The published inside pipe diameter.

3.5

dynamic performance

A general expression for the relationship between the volume or mass registered by a meter and the true value after a change in flow rate, flowing temperature or flowing pressure.

¹ International Electrotechnical Commission, 3, rue de Varembé, P.O. Box 131, CH-1211, Geneva 20, Switzerland, www.iec.ch.