

# **Manual of Petroleum Measurement Standards Chapter 22.5**

## **Testing Protocols—Flow Computational Device**

FIRST EDITION, MARCH 2024



American  
Petroleum  
Institute

## Special Notes

API publications necessarily address problems of a general nature. With respect to particular circumstances, local, state, and federal laws and regulations should be reviewed. The use of API publications is voluntary. In some cases, third parties or authorities having jurisdiction may choose to incorporate API standards by reference and may mandate compliance.

Neither API nor any of API's employees, subcontractors, consultants, committees, or other assignees make any warranty or representation, either express or implied, with respect to the accuracy, completeness, or usefulness of the information contained herein, or assume any liability or responsibility for any use, or the results of such use, of any information or process disclosed in this publication. Neither API nor any of API's employees, subcontractors, consultants, or other assignees represent that use of this publication would not infringe upon privately owned rights.

API publications may be used by anyone desiring to do so. Every effort has been made by the Institute to ensure the accuracy and reliability of the data contained in them; however, the Institute makes no representation, warranty, or guarantee in connection with this publication and hereby expressly disclaims any liability or responsibility for loss or damage resulting from its use or for the violation of any authorities having jurisdiction with which this publication may conflict.

API publications are published to facilitate the broad availability of proven, sound engineering and operating practices. These publications are not intended to obviate the need for applying sound engineering judgment regarding when and where these publications should be used. The formulation and publication of API publications is not intended in any way to inhibit anyone from using any other practices.

Any manufacturer marking equipment or materials in conformance with the marking requirements of an API standard is solely responsible for complying with all the applicable requirements of that standard. API does not represent, warrant, or guarantee that such products do in fact conform to the applicable API standard.

All rights reserved. No part of this work may be reproduced, translated, stored in a retrieval system, or transmitted by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from the publisher. Contact the Publisher, API Publishing Services, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001-5571.

## Foreword

Nothing contained in any API publication is to be construed as granting any right, by implication or otherwise, for the manufacture, sale, or use of any method, apparatus, or product covered by letters patent. Neither should anything contained in the publication be construed as insuring anyone against liability for infringement of letters patent.

The verbal forms used to express the provisions in this document are as follows.

Shall: As used in a standard, “shall” denotes a minimum requirement to conform to the standard.

Should: As used in a standard, “should” denotes a recommendation or that which is advised but not required to conform to the standard.

May: As used in a standard, “may” denotes a course of action permissible within the limits of a standard.

Can: As used in a standard, “can” denotes a statement of possibility or capability.

This document was produced under API standardization procedures that ensure appropriate notification and participation in the developmental process and is designated as an API standard. Questions concerning the interpretation of the content of this publication or comments and questions concerning the procedures under which this publication was developed should be directed in writing to the Director of Standards, American Petroleum Institute, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001. Requests for permission to reproduce or translate all or any part of the material published herein should also be addressed to the director.

Generally, API standards are reviewed and revised, reaffirmed, or withdrawn at least every five years. A one-time extension of up to two years may be added to this review cycle. Status of the publication can be ascertained from the API Standards Department, telephone (202) 682-8000. A catalog of API publications and materials is published annually by API, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001.

Suggested revisions are invited and should be submitted to the Standards Department, API, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001, [standards@api.org](mailto:standards@api.org).



# Contents

	Page
<b>1</b> Scope.....	1
<b>2</b> Normative References .....	1
<b>3</b> Terms, Definitions, and Symbols .....	1
<b>3.1</b> Terms and Definitions .....	1
<b>3.2</b> Symbols .....	2
<b>4</b> Applicability .....	2
<b>5</b> Parameter Variations Affecting Device Performance .....	2
<b>5.1</b> Dynamic Calculation .....	2
<b>6</b> Mandatory Tests.....	2
<b>6.1</b> Test Conditions .....	2
<b>6.2</b> Test Installation .....	2
<b>6.3</b> Test Results .....	4
<b>6.4</b> Baseline (Ideal Condition) Testing .....	4
<b>6.5</b> Dynamic Flow Condition Testing.....	4
<b>6.6</b> Special Testing.....	4
<b>6.7</b> Testing Documentation .....	5
<b>6.8</b> Testing Procedure .....	5
<b>7</b> Test Facility Requirements.....	9
<b>7.1</b> Audit Process.....	9
<b>7.2</b> Laboratory/Facility Qualification.....	9
<b>7.3</b> Validity/Precision of Laboratory/Facility Results .....	9
<b>8</b> Uncertainty Analysis and Calculation.....	9
<b>8.1</b> Types of Uncertainty Calculations .....	9
<b>8.2</b> How to Calculate Uncertainty .....	10
<b>8.3</b> Presentation of Uncertainty Test Report .....	10
<b>9</b> Test Report .....	10
<b>9.1</b> General .....	10
<b>9.2</b> Static Testing.....	11
<b>9.3</b> Volume Flowrate at User Defined Base Conditions—Dynamic Testing .....	13
Annex A (normative) Test Points for Static Testing of AGA Report Number 8 and AGA Report Number 5 or GPA 2172, <i>Calculations</i> .....	15
Annex B (normative) Test Points for Static Testing of API <i>MPMS</i> Chapter 14.3, <i>Calculations</i> .....	18
Annex C (normative) Flow Patterns and Parameters for Dynamic Testing .....	20
Bibliography.....	28

# Contents

Page

## Figures

<b>1</b>	Flow Computation Device Test Installation .....	3
<b>2</b>	Block Diagram of Test Set-up Algorithm Verification Process.....	8
<b>C.1</b>	Example of an Over-damped I/P Output.....	21
<b>C.2</b>	Example of an Under-damped I/P Output.....	22
<b>C.3</b>	I/P with Poor Output Control .....	23
<b>C.4</b>	I/P with Good Output Control .....	24
<b>C.5</b>	Example of One-Hour Differential Pressure Trend with a Five-Minute Offset.....	26
<b>C.6</b>	Late Trend Last Five Minutes Matches First Five Minutes of the On-Time Trend .....	27

## Tables

<b>A.1</b>	Compositions To Be Tested (Johansen and Johnson [18]).....	17
------------	--	----

## Introduction

This document defines the testing and reporting protocols for devices used to calculate flowrates or volumes from orifice meters measuring hydrocarbon fluids. This protocol is designed to supply industry with a baseline of capabilities of these devices under defined operating conditions in this document. The objectives of this testing protocol are to:

- 1) Ensure that the user of any flow computation device knows the performance characteristics of the device over an applicable range as defined by tests.
- 2) Facilitate both the understanding and the introduction of new technologies.
- 3) Provide a standardized vehicle for validating manufacturers' performance specifications.
- 4) Provide information about relative performance characteristics of the flow computation devices under standardized testing protocol.
- 5) Quantify the uncertainty of these devices and define the operating and installation conditions for which the stated uncertainties apply.

To accomplish these objectives, the testing protocol defines the test limits for operating conditions of the devices, and the requirements of the facility or facilities to perform the tests. This document encompasses any device capable of calculating flowrates as measured by an orifice meter.



# Testing Protocols—Flow Computational Device

## 1 Scope

This API *MPMS* chapter is intended to be a standardized guideline to determine and document the fidelity of implementation of the API *MPMS* calculations standards. As with all standards developed under Chapter 22.1, this standard includes a listing of parameters affecting the performance of the devices, a description of the tests required, requirements for the test facility, a data reporting format, and an uncertainty determination methodology.

The field of application for this testing protocol is limited to devices that are used in the orifice-based measurement of hydrocarbon fluids in the petroleum, energy, and petrochemical industries.

## 2 Normative References

There are no normative references for this document.

## 3 Terms, Definitions, and Symbols

### 3.1 Terms and Definitions

For the purposes of this document, the following terms and definitions apply.

#### 3.1.1

##### **accuracy**

The extent to which the results of a calculation or the readings of an instrument approach the true value.

#### 3.1.2

##### **calibration span**

The difference between the calibrated maximum and minimum range limits

#### 3.1.3

##### **certified equipment**

Equipment whose performance is traceable to primary standards maintained by an internationally recognized standards organization, such as the National Institute of Standards and Technology, and that has been provided with documentation (Certificate of Conformance) stating the traceability.

#### 3.1.4

##### **flow computation device**

An arithmetic processing unit with associated memory that accepts electrically converted signals representing input variables from a measurement system and performs calculation for the purpose of providing flow rate and total quantity data.

#### 3.1.5

##### **calculation version identifier**

An auditable (see API *MPMS* Chapter 21.1) unique identifier used to distinguish any changes to the calculation methodology of a flow computation device.

#### 3.1.6

##### **non-ideal testing**

Typically defined by field conditions during a system start-up (FAT/SAT).

#### 3.1.7

##### **quantity calculation period (QCP)**

The period over which the calculated total quantity is to be integrated.