

# **Manual of Petroleum Measurement Standards Chapter 9.3**

## **Standard Test Method for Density, Relative Density, and API Gravity of Crude Petroleum and Liquid Petroleum Products by Thermohydrometer Method**

FIFTH EDITION, NOVEMBER 2024



American  
Petroleum  
Institute

## Special Notes

API publications necessarily address problems of a general nature. With respect to particular circumstances, local, state, and federal laws and regulations should be reviewed. The use of API publications is voluntary. In some cases, third parties or authorities having jurisdiction may choose to incorporate API standards by reference and may mandate compliance.

Neither API nor any of API's employees, subcontractors, consultants, committees, or other assignees make any warranty or representation, either express or implied, with respect to the accuracy, completeness, or usefulness of the information contained herein, or assume any liability or responsibility for any use, or the results of such use, of any information or process disclosed in this publication. Neither API nor any of API's employees, subcontractors, consultants, or other assignees represent that use of this publication would not infringe upon privately owned rights.

Users of this standard should not rely exclusively on the information contained in this document. Sound business, scientific, engineering, and safety judgment should be used in employing the information contained herein.

API is not undertaking to meet the duties of employers, manufacturers, or suppliers to warn and properly train and equip their employees, and others exposed, concerning health and safety risks and precautions, nor undertaking their obligations to comply with authorities having jurisdiction.

API publications may be used by anyone desiring to do so. Every effort has been made by the Institute to ensure the accuracy and reliability of the data contained in them; however, the Institute makes no representation, warranty, or guarantee in connection with this publication and hereby expressly disclaims any liability or responsibility for loss or damage resulting from its use or for the violation of any authorities having jurisdiction with which this publication may conflict.

API publications are published to facilitate the broad availability of proven, sound engineering and operating practices. These publications are not intended to obviate the need for applying sound engineering judgment regarding when and where these publications should be utilized. The formulation and publication of API publications is not intended in any way to inhibit anyone from using any other practices.

Any manufacturer marking equipment or materials in conformance with the marking requirements of an API standard is solely responsible for complying with all the applicable requirements of that standard. API does not represent, warrant, or guarantee that such products do in fact conform to the applicable API standard.

All rights reserved. No part of this work may be reproduced, translated, stored in a retrieval system, or transmitted by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from the publisher. Contact the publisher, API Publishing Services, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001.

## Foreword

Nothing contained in any API publication is to be construed as granting any right, by implication or otherwise, for the manufacture, sale, or use of any method, apparatus, or product covered by letters patent. Neither should anything contained in the publication be construed as insuring anyone against liability for infringement of letters patent.

The verbal forms used to express the provisions in this document are as follows.

Shall: As used in a standard, “shall” denotes a minimum requirement in order to conform to the standard.

Should: As used in a standard, “should” denotes a recommendation or that which is advised but not required in order to conform to the standard.

May: As used in a standard, “may” denotes a course of action permissible within the limits of a standard.

Can: As used in a standard, “can” denotes a statement of possibility or capability.

This document was produced under API standardization procedures that ensure appropriate notification and participation in the developmental process and is designated as an API standard. Questions concerning the interpretation of the content of this publication or comments and questions concerning the procedures under which this publication was developed should be directed in writing to the Director of Standards, American Petroleum Institute, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001. Requests for permission to reproduce or translate all or any part of the material published herein should also be addressed to the director.

Generally, API standards are reviewed and revised, reaffirmed, or withdrawn at least every five years. A one-time extension of up to two years may be added to this review cycle. Status of the publication can be ascertained from the API Standards Department, telephone (202) 682-8000. A catalog of API publications and materials is published annually by API, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001.

Suggested revisions are invited and should be submitted to the Standards Department, API, 200 Massachusetts Avenue, NW, Suite 1100, Washington, DC 20001, [standards@api.org](mailto:standards@api.org).



# Contents

	<b>Page</b>
1 Scope .....	1
2 Referenced Documents .....	1
3 Terminology .....	2
4 Summary of Test Method.....	3
5 Significance and Use.....	4
6 Apparatus.....	4
7 Sampling, Test Specimens, and Test Units .....	8
8 Apparatus Verification or Certification .....	8
9 Procedure.....	8
10 Calculation .....	13
11 Reports.....	14
12 Precision and Bias.....	14
13 Keywords .....	15
Annex A1—Equipment Validation as to Fit for Use at Time of Test.....	15
Annex A2—Performance Verification of Density and Temperature Measuring Instruments.....	16
Annex A3—Calibration of Density and Temperature Measuring Instruments.....	17
Summary of Changes .....	18

## Figures

1 Typical Thermohydrometer Designs .....	4
2 Typical Separate Density and Temperature Measuring Devices .....	7
3 Typical Single Density and Temperature Measuring Device .....	7
4 Hydrometer Reading for Transparent Fluids.....	10
5 Hydrometer Reading for Opaque Fluids .....	10
6 Hydrometer and Liquid-in-Glass Thermometer Placement (Typical).....	11
7 Thermohydrometer and Liquid-in-Glass Thermometer Placement (Typical).....	11
8 Hydrometer and Digital Contact Thermometer (Typical).....	11
9 Thermohydrometer and Digital Contact Thermometer Placement (Typical) .....	11
10 Hydrometer Scale Reading for Transparent Liquids (Typical).....	12
11 Hydrometer Scale Reading for Opaque Fluids (Typical).....	12

## Tables

1 Density Thermohydrometers .....	5
2 API Gravity Thermohydrometers .....	6
3 Thermometers (Liquid-in-Glass).....	7
4 Limiting Conditions and Test Temperatures .....	8





Designation: D6822 – 24



Manual of Petroleum Measurement Standards (MPMS), Chapter 9.3

# Standard Test Method for Density, Relative Density, and API Gravity of Crude Petroleum and Liquid Petroleum Products by Thermohydrometer Method<sup>1</sup>

This standard is issued under the fixed designation D6822; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reappraisal.

## 1. Scope\*

1.1 This test method covers the determination, using a glass thermohydrometer and/or a separate (alternative) temperature measuring device in conjunction with a series of calculations, of the density, relative density, or API gravity of crude petroleum, petroleum products, or mixtures of petroleum and nonpetroleum products normally handled as liquids and having Reid vapor pressures (Test Method **D323**) of 101.325 kPa (14.696 psia) or less. Values are determined at existing temperatures and corrected to 15 °C (or 60 °F) by means of a series of calculations and international standard tables. Alternative temperature measuring devices shall be either the appropriate liquid-in-glass ASTM thermometers meeting either Specifications **E1** or **E2251** or digital contact thermometers (Guides **E2877** and **D8164**).

1.2 The initial thermohydrometer readings obtained are uncorrected hydrometer readings and not density measurements. Readings are measured on a thermohydrometer at either the reference temperature or at another convenient temperature, and readings are corrected for the meniscus effect, the thermal glass expansion effect, alternate calibration temperature effects and to the reference temperature by means of calculations and Adjunct to **D1250** Guide for Use of the Petroleum Measurement Tables (API *MPMS* Chapter 11.1). Values measured (determined) at other than the reference temperature being hydrometer/thermohydrometer readings and not density measurements

1.3 Readings determined as density, relative density, or API gravity can be converted to equivalent values in the other units

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee **D02** on Petroleum Products, Liquid Fuels, and Lubricants and the API Committee on Petroleum Measurement, and is the direct responsibility of Subcommittee **D02.02** /COMQ, the joint ASTM-API Committee on Hydrocarbon Measurement for Custody Transfer (Joint ASTM-API). This test method has been approved by the sponsoring committees and accepted by the Cooperating Societies in accordance with established procedures.

Current edition approved July 1, 2024. Published November 2024. Originally approved in 2002. Last previous edition approved in 2023 as D6822 – 23. DOI: 10.1520/D6822-24.

or alternate reference temperatures by means of Interconversion Procedures (API *MPMS* Chapter 11.5) or Adjunct to **D1250** Guide for Use of the Petroleum Measurement Tables (API *MPMS* Chapter 11.1), or both, or tables as applicable.

1.4 The initial thermohydrometer reading (including temperature) shall be recorded before performing any calculations. The calculations required in **9.1.5** and Section **10** shall be applied to the initial thermohydrometer reading with observations and results reported as required by Section **11** prior to use in a subsequent calculation procedure (measurement ticket calculation, meter factor calculation, or base prover volume determination).

1.5 **Annex A1 – Annex A3** contain procedures for determining fit-for-service, verifying, or certifying the equipment for this test method.

1.6 The values stated in SI units are to be regarded as standard.

1.6.1 *Exception*—The values given in parentheses (inch-pound) are for information only.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.8 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

### 2.1 *ASTM Standards*:<sup>2</sup>

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [www.astm.org/contact](http://www.astm.org/contact). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

\*A Summary of Changes section appears at the end of this standard