

Manual of Petroleum Measurement Standards Chapter 10.4

Determination of Water and/or Sediment in Crude Oil by the Centrifuge Method (Field Procedure)

FOURTH EDITION, OCTOBER 2013

ERRATA, MARCH, 2015



AMERICAN PETROLEUM INSTITUTE

Special Notes

API publications necessarily address problems of a general nature. With respect to particular circumstances, local, state, and federal laws and regulations should be reviewed.

Neither API nor any of API's employees, subcontractors, consultants, committees, or other assignees make any warranty or representation, either express or implied, with respect to the accuracy, completeness, or usefulness of the information contained herein, or assume any liability or responsibility for any use, or the results of such use, of any information or process disclosed in this publication. Neither API nor any of API's employees, subcontractors, consultants, or other assignees represent that use of this publication would not infringe upon privately owned rights.

API publications may be used by anyone desiring to do so. Every effort has been made by the Institute to assure the accuracy and reliability of the data contained in them; however, the Institute makes no representation, warranty, or guarantee in connection with this publication and hereby expressly disclaims any liability or responsibility for loss or damage resulting from its use or for the violation of any authorities having jurisdiction with which this publication may conflict.

API publications are published to facilitate the broad availability of proven, sound engineering and operating practices. These publications are not intended to obviate the need for applying sound engineering judgment regarding when and where these publications should be utilized. The formulation and publication of API publications is not intended in any way to inhibit anyone from using any other practices.

Any manufacturer marking equipment or materials in conformance with the marking requirements of an API standard is solely responsible for complying with all the applicable requirements of that standard. API does not represent, warrant, or guarantee that such products do in fact conform to the applicable API standard.

Users of this standard should not rely exclusively on the information contained in this document. Sound business, scientific, engineering, and safety judgment should be used in employing the information contained herein.

API is not undertaking to meet the duties of employers, manufacturers, or suppliers to warn and properly train and equip their employees, and others exposed, concerning health and safety risks and precautions, nor undertaking their obligations to comply with authorities having jurisdiction.

All rights reserved. No part of this work may be reproduced, translated, stored in a retrieval system, or transmitted by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from the publisher. Contact the Publisher, API Publishing Services, 1220 L Street, NW, Washington, DC 20005.

Copyright © 2013 American Petroleum Institute

Foreword

Nothing contained in any API publication is to be construed as granting any right, by implication or otherwise, for the manufacture, sale, or use of any method, apparatus, or product covered by letters patent. Neither should anything contained in the publication be construed as insuring anyone against liability for infringement of letters patent.

Shall: As used in a standard, “shall” denotes a minimum requirement in order to conform to the specification.

Should: As used in a standard, “should” denotes a recommendation or that which is advised but not required in order to conform to the specification.

This document was produced under API standardization procedures that ensure appropriate notification and participation in the developmental process and is designated as an API standard. Questions concerning the interpretation of the content of this publication or comments and questions concerning the procedures under which this publication was developed should be directed in writing to the Director of Standards, American Petroleum Institute, 1220 L Street, NW, Washington, DC 20005. Requests for permission to reproduce or translate all or any part of the material published herein should also be addressed to the director.

Generally, API standards are reviewed and revised, reaffirmed, or withdrawn at least every five years. A one-time extension of up to two years may be added to this review cycle. Status of the publication can be ascertained from the API Standards Department, telephone (202) 682-8000. A catalog of API publications and materials is published annually by API, 1220 L Street, NW, Washington, DC 20005.

Suggested revisions are invited and should be submitted to the Standards Department, API, 1220 L Street, NW, Washington, DC 20005, standards@api.org.

Contents

	Page
1 Scope	1
2 Normative References	1
3 Terms and Definitions	2
4 Significance and Use	2
5 Test Solvents	3
5.1 General	3
5.2 Use of Solvents	3
6 Demulsifier	3
7 Apparatus	4
7.1 Centrifuge	4
7.2 Centrifuge Tubes	5
7.3 Preheater	5
7.4 Sample Thermometer	5
8 Sampling	5
9 Procedure—Determination of Water and Sediment	5
9.1 General	5
9.2 Procedure	8
10 Calculation and Reporting—Water and Sediment	9
10.1 General	9
10.2 100-mL Tubes	9
10.3 200-part Tubes	9
11 Procedure—Determination of Sediment Only	10
11.1 General	10
11.2 Procedure	10
12 Calculation and Reporting—Sediment Only	11
12.1 General	11
12.2 100-mL Tubes	11
12.3 200-part Tubes	11
13 Precision	11
14 Bias	11
Annex A (informative) Precautionary Information	12
Annex B (informative) Demulsifiers and Water Saturation of Solvents	15
Annex C (normative) Specifications for Centrifuge Tubes	16
Annex D (informative) Procedure Outline—Determination of Water and Sediment	21
Annex E (informative) Procedure Outline—Determination of Sediment Only	22
Bibliography	23

Contents

Page

Figures

1	Reading a 100-mL Centrifuge Tube	6
2	Reading a 200-Part Centrifuge Tube	7
C-1	8-in. (203-mm) Cone-Shaped Centrifuge Tube	17
C-2	6-in. (167-mm) Cone-Shaped Centrifuge Tube	19

Tables

1	Rotation Speeds Necessary to Produce a Relative Centrifugal Force of 500 g for Centrifuges of Various Diameters of Swing	4
2	Procedure for Reading a 100-mL Cone-Shaped Tube	6
3	Procedure for Reading a 200-Part Cone-Shaped Tube	7
4	Expression of Results for 100-mL Tubes and 200-Part Tubes	9
C-1	Minimum Graduation Requirements and Maximum Calibration Tolerances for 8-in. (203-mm) Cone-Shaped Tubes	18
C-2	Minimum Graduation Requirements and Maximum Calibration Tolerances for 6-in. (167-mm) Cone-Shaped Tubes	20

Determination of Water and/or Sediment in Crude Oil by the Centrifuge Method (Field Procedure)

1 Scope

1.1 This section describes the field centrifuge method for determining both water and sediment or sediment only in crude oil. This method may not always produce the most accurate results, but it is considered the most practical method for field determination of water and sediment. This method may also be used for field determination of sediment. When a higher degree of accuracy is required, the laboratory procedure described in API MPMS Ch. 10.3, *Standard Test Method for Water and Sediment in Crude Oil by the Centrifuge Method (Laboratory Procedure)* (ASTM D4007); API MPMS Ch. 10.2, *Standard Test Method for Water in Crude Oil by Distillation* (ASTM D4006); or API MPMS Ch. 10.9, *Standard Test Method for Water in Crude Oils by Coulometric Karl Fischer Titration* (ASTM D4928); and the procedure described in API MPMS Ch. 10.1, *Standard Test Method for Sediment in Crude Oils and Fuel Oils by the Extraction Method* (ASTM D473) or API MPMS Ch. 10.8, *Standard Test Method for Sediment in Crude Oil by Membrane Filtration* (ASTM D4807) should be used.

1.2 The requirements of this standard may involve hazardous materials, operations, and equipment. This standard does not purport to address all of the safety problems associated with its use. The user of this standard has the responsibility for establishing appropriate safety and health practices and determining the applicability of regulatory limitations prior to use.

2 Normative References

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

API MPMS Chapter 8.1, *Manual Sampling of Petroleum and Petroleum Products*

API MPMS Chapter 8.2, *Automatic Sampling of Petroleum and Petroleum Products*

API MPMS Chapter 8.3, *Standard Practice for Mixing and Handling of Liquid Samples of Petroleum and Petroleum Products (includes Errata 1 dated March 1996)*

API MPMS Chapter 10.1, *Standard Test Method for Sediment in Crude Oils and Fuel Oils by the Extraction Method*

API MPMS Chapter 10.2, *Standard Test Method for Water in Crude Oil by Distillation*

API MPMS Chapter 10.3, *Standard Test Method for Water and Sediment in Crude Oil by the Centrifuge Method (Laboratory Procedure)*

API MPMS Chapter 10.8, *Standard Test Method for Sediment in Crude Oil by Membrane Filtration*

API MPMS Chapter 10.9, *Standard Test Method for Water in Crude Oils by Coulometric Karl Fischer Titration*

ASTM D473 ¹, *Standard Test Method for Sediment in Crude Oils and Fuel Oils by the Extraction Method*

ASTM D3699, *Standard Specification for Kerosine*

ASTM D4006, *Standard Test Method for Water in Crude Oil by Distillation*

¹ ASTM International, 100 Barr Harbor Drive, West Conshohocken, Pennsylvania 19428, www.astm.org.