

Manual of Petroleum Measurement Standards Chapter 10—Sediment and Water

Section 4—Determination of Water and/or Sediment in Crude Oil by the Centrifuge Method (Field Procedure)

THIRD EDITION, DECEMBER 1999

REAFFIRMED, SEPTEMBER 2010



AMERICAN PETROLEUM INSTITUTE

Manual of Petroleum Measurement Standards Chapter 10—Sediment and Water

**Section 4—Determination of Water and/or
Sediment in Crude Oil by the
Centrifuge Method (Field Procedure)**

Measurement Coordination

THIRD EDITION, DECEMBER 1999

REAFFIRMED, SEPTEMBER 2010



AMERICAN PETROLEUM INSTITUTE

SPECIAL NOTES

API publications necessarily address problems of a general nature. With respect to particular circumstances, local, state, and federal laws and regulations should be reviewed.

API is not undertaking to meet the duties of employers, manufacturers, or suppliers to warn and properly train and equip their employees, and others exposed, concerning health and safety risks and precautions, nor undertaking their obligations under local, state, or federal laws.

Information concerning safety and health risks and proper precautions with respect to particular materials and conditions should be obtained from the employer, the manufacturer or supplier of that material, or the material safety data sheet.

Nothing contained in any API publication is to be construed as granting any right, by implication or otherwise, for the manufacture, sale, or use of any method, apparatus, or product covered by letters patent. Neither should anything contained in the publication be construed as insuring anyone against liability for infringement of letters patent.

Generally, API standards are reviewed and revised, reaffirmed, or withdrawn at least every five years. Sometimes a one-time extension of up to two years will be added to this review cycle. This publication will no longer be in effect five years after its publication date as an operative API standard or, where an extension has been granted, upon republication. Status of the publication can be ascertained from API Measurement Coordination [telephone (202) 682-8000]. A catalog of API publications and materials is published annually and updated quarterly by API, 1220 L Street, N.W., Washington, D.C. 20005.

This document was produced under API standardization procedures that ensure appropriate notification and participation in the developmental process and is designated as an API standard. Questions concerning the interpretation of the content of this standard or comments and questions concerning the procedures under which this standard was developed should be directed in writing to the Standardization Manager, American Petroleum Institute, 1220 L Street, N.W., Washington, D.C. 20005. Requests for permission to reproduce or translate all or any part of the material published herein should also be addressed to the director.

API standards are published to facilitate the broad availability of proven, sound engineering and operating practices. These standards are not intended to obviate the need for applying sound engineering judgment regarding when and where these standards should be utilized. The formulation and publication of API standards is not intended in any way to inhibit anyone from using any other practices.

Any manufacturer marking equipment or materials in conformance with the marking requirements of an API standard is solely responsible for complying with all the applicable requirements of that standard. API does not represent, warrant, or guarantee that such products do in fact conform to the applicable API standard.

All rights reserved. No part of this work may be reproduced, stored in a retrieval system, or transmitted by any means, electronic, mechanical, photocopying, recording, or otherwise, without prior written permission from the publisher. Contact the Publisher, API Publishing Services, 1220 L Street, N.W., Washington, D.C. 20005.

FOREWORD

The American Petroleum Institute, the (British) Institute of Petroleum (IP), and corresponding bodies of other nations have worked through Technical Committee 28 of the International Organization for Standardization (ISO) to produce several international standards that relate directly to the *API Manual of Petroleum Measurement Standards*.

Section 4 of Chapter 10 discusses the field centrifuge test method for determining water and/or sediment in crude oil. Section 4 is primarily intended for use within the United States and has not been approved by IP or ISO. This edition of Section 4 of Chapter 10 supersedes all previous editions of this standard.

Section 1 of Chapter 10 describes the determination of sediment in crude oils and fuel oils by the extraction method. Section 2 discusses the determination of water in crude oil by the distillation method. Section 3 describes the determination of sediment and water by a laboratory centrifuge procedure.

This standard may involve hazardous materials, operations, and equipment. This standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

API publications may be used by anyone desiring to do so. Every effort has been made by the Institute to assure the accuracy and reliability of the data contained in them; however, the Institute makes no representation, warranty, or guarantee in connection with this publication and hereby expressly disclaims any liability or responsibility for loss or damage resulting from its use or for the violation of any federal, state, or municipal regulation with which this publication may conflict.

Suggested revisions are invited and should be submitted to Measurement Coordination, American Petroleum Institute, 1220 L Street, N.W., Washington, D.C. 20005.

CONTENTS

	Page
10.4.1 SCOPE AND FIELD OF APPLICATION	1
10.4.2 REFERENCES	1
10.4.3 DEFINITIONS	1
10.4.4 SIGNIFICANCE AND USE	1
10.4.5 TEST SOLVENTS	2
10.4.5.1 Use of Solvents	2
10.4.6 DEMULSIFIER	2
10.4.7 APPARATUS	2
10.4.7.1 Centrifuge	2
10.4.7.2 Centrifuge Tubes	3
10.4.7.3 Preheater	3
10.4.7.4 Sample Thermometer	3
10.4.8 SAMPLING	3
10.4.9 PROCEDURE—WATER AND SEDIMENT DETERMINATION	3
10.4.10 CALCULATION AND REPORTING	6
10.4.11 DETERMINATION OF SEDIMENT ONLY	6
10.4.11.1 Procedure	6
10.4.12 CALCULATION AND REPORTING—SEDIMENT ONLY	7
10.4.13 PRECISION	8
10.4.14 BIAS	8
APPENDIX A—PRECAUTIONARY INFORMATION	9
APPENDIX B—DEMULSIFIERS AND WATER SATURATION OF SOLVENTS	11
APPENDIX C—SPECIFICATIONS FOR CENTRIFUGE TUBES	13
APPENDIX D—PROCEDURE OUTLINE—WATER AND SEDIMENT DETERMINATION	17
APPENDIX E—PROCEDURE OUTLINE—SEDIMENT (ONLY) DETERMINATION	19
Figures	
1—Reading a 100-milliliter Centrifuge Tube	4
2—Reading a 200-Part Centrifuge Tube	5
C-1—8-inch (203-millimeter) Cone-Shaped Centrifuge Tube	13
C-2—6-inch (167-millimeter) Cone-Shaped Centrifuge Tube	14

CONTENTS

Page

Tables

1—Rotation Speeds Necessary to Produce a Relative Centrifugal Force of 500 for Centrifuges of Various Diameters of Swing	3
2—Procedure for Reading a 100-milliliter Cone-Shaped Tube	4
3—Procedure for Reading a 200-Part Cone-Shaped Tube	5
4—Expression of Results	6
C-1—Minimum Graduation Requirements and Maximum Calibration Tolerances for 8-inch (203-millimeter) Cone-Shaped Tubes	14
C-2—Minimum Graduation Requirements and Maximum Calibration Tolerances for 6-inch (167-millimeter) Cone-Shaped Tubes	15

Chapter 10—Sediment and Water

SECTION 4—DETERMINATION OF WATER AND/OR SEDIMENT IN CRUDE OIL BY THE CENTRIFUGE METHOD (FIELD PROCEDURE)

10.4.1 Scope and Field of Application

10.4.1.1 This section describes the field centrifuge method for determining both water and sediment or sediment only in crude oil. This method may not always produce the most accurate results, but it is considered the most practical method for field determination of water and/or sediment. When a higher degree of accuracy is required, the laboratory procedure described in Chapter 10.2, “Determination of Water in Crude Oil by Distillation” (ASTM D 4006), or Chapter 10.9, “Standard Test Method for Water in Crude Oils by Coulometric Karl Fischer” (ASTM D 4928), and the procedure described in Chapter 10.1, “Determination of Sediment in Crude Oils and Fuel Oils by the Extraction Method” (ASTM D 473) or Chapter 10.8, “Standard Test Method for Sediment in Crude Oil by Membrane Filtration” (ASTM D 4807) should be used.

Note: Water by Distillation, and Sediment by Extraction, are considered the referee methods for determining water and sediment in crude oils.

10.4.1.2 The requirements of this standard may involve hazardous materials, operations, and equipment. This standard does not purport to address all of the safety problems associated with its use. The user of this standard has the responsibility for establishing appropriate safety and health practices and determining the applicability of regulatory limitations prior to use.

10.4.2 References

API

Manual of Petroleum Measurement Standards

Chapter 8, “Sampling Petroleum and Petroleum Products”

Chapter 10, “Sediment and Water”

ACGIH¹

Threshold Limit Values and Biological Exposure Indices for 1987 – 1988

ASTM²

D 473 *Test Method for Sediment in Crude Oils and Fuel Oils by the Extraction Method*

D 3699 *Specification for Kerosine*

D 4006 *Test Method for Water in Crude Oil by Distillation*

D 4057 *Practice for Manual Sampling of Petroleum and Petroleum Products*

D 4177 *Method for Automatic Sampling of Petroleum and Petroleum Products*

D 4928 *Standard Test Method for Water in Crude Oils By Coulometric Karl Fischer Titration (Karl Fischer) Titration*

E 542 *Practice for Calibration of Volumetric Ware*

OSHA³

29 *CFR* Subpart Z, “Toxic and Hazardous Substances,” Section 1910.1000

10.4.3 Definitions

10.4.3.1 sediment: A material that normally exists in crude oils as extremely fine, well dispersed solids that originate in the reservoir from which the crude came, or in drilling fluids used to drill the wells, which may take the form of sand, clay, shale, or rock particles. Other sediments such as scale can be picked up from tubing, pipe, tanks, and other production and transportation equipment. Sediment, as defined in this standard, must have a density greater than water and will not float at test conditions.

10.4.3.2 demulsifier-stock solution: A mixture of demulsifier and solvent in a fixed ratio (e.g. 25% – 75%) which is added to the test solvent to form a solution to cause a clear and distinct separation between the oil and water interface.

10.4.3.3 solvent-solution: A mixture of any one of the approved solvents and a demulsifier, or a demulsifier-stock solution which is used as a diluent to help the separation of water and sediment from the sample being tested.

10.4.3.4 solvent: Any one of the approved liquids capable of dissolving or dispersing the water and sediment in the sample being tested.

10.4.4 Significance and Use

A determination of water and/or sediment content is required to accurately determine the net volumes of crude oil

¹American Conference of Governmental Industrial Hygienists, 1330 Kemper Meadow Drive, Cincinnati, Ohio 45240.

²American Society for Testing and Materials, 100 Barr Harbor Drive, West Conshohocken, Pennsylvania 19428-2959.

³Occupational Safety and Health Administration, U.S. Department of Labor. *The Code of Federal Regulations* is available from the Government Printing Office, Washington, D.C. 20402.